

Test Report

Number: GZHT02340911


Report Ref:	GZHT02340911		
Date Received:	Sep 21, 2020	Date Issued:	Sep 24, 2020

Company Name:	SHANDONG TELLER TEXTILE CO.,LTD		
Address:	TAOYUAN INDUSTRIAL PARK WULIAN COUNTRY,RIZHAO CITY SHANGDONG PROVINCE,CHINA		
Contact Name:	CHERRY		

Sample Description:			
The Following Sample Was Submitted And Identified By/On Behalf Of The Applicant As:			
Sample Name	:	Isolation Gown	
No. Of Sample	:	One (6 Pieces)	
Size(s) Range	:	-	
Color	:	Blue	
Classifications Claimed	:	Level 3	
End Uses	:	Isolation Gown	
Standard	:	ANSI/AAMI PB70:2012	
Manufacturer	:	SHANDONG TELLER TEXTILE CO.,LTD	
Date received/ Date Test Started	:	Sep 21, 2020	
Date Final Information Confirmed	:	-	
Ref	:	Type No.: TLM 005	

Test was conducted on specific items, at our client's request.

Prepared And Checked By:
For Intertek Testing Services Shenzhen Ltd. Guangzhou Branch



Lin Lin
General Manager



Ula / hilaryxu

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch

深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E801,
No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China

中国广州经济技术开发区科学城彩频路7号之二第1-8层02房、01房
101、

E201, E301, E401, E501, E601, E701, E801

Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663

Room 03, Hengyun Building, 235 Kaifa Ave., Guangzhou

Economic & Technological Development District, Guangzhou, China

中国广州经济技术开发区开发大道235号恒运大厦3楼

Tel: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730



Original Sample Photo



Prepared And Checked By:
For Intertek Testing Services Shenzhen Ltd. Guangzhou Branch

Lin Lin
General Manager

Ula / hilaryxu



Intertek Testing Services Shenzhen Ltd. Guangzhou Branch

深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E801,
No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China

中国广州经济技术开发区科学城彩频路7号之二第1-8层02房、01房、

101、E201、E301、E401、E501、E601、E701、E801

Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663

Room 03, Hengyun Building, 235 Kaifa Ave., Guangzhou

Economic & Technological Development District, Guangzhou, China

中国广州经济技术开发区开发大道235号恒运大厦3楼

Tel: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730



Test Report

Number: GZHT02340911

Summary of Testing:


With reference to following standard:

- ANSI/AAMI PB70:2012 Liquid Barrier Performance and Classification of Protective apparel and Drapes intended for Use in Health Care Facilities Level 3

Tested Components in The Submitted Samples Were Found To Comply With The Level 3 Requirements of ANSI/AAMI PB70:2012 with respect to Water Resistance: Impact Penetration and Hydrostatic Pressure tests.

Remark: Tested and judged on client's requirement without considering the AQL of 4 %/RQL of 20 % levels for the sampling plans employed.

Prepared And Checked By:
For Intertek Testing Services Shenzhen Ltd. Guangzhou Branch



Lin Lin
General Manager



Ula / hilaryxu

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch

深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E801,
No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China

中国广州经济技术开发区科学城彩频路7号之二第1-8层02房、01房、101、

E201、E301、E401、E501、E601、E701、E801

Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E801,
Hengyun Building, 235 Kaifa Ave., Guangzhou

Economic & Technological Development District, Guangzhou, China

中国广州经济技术开发区开发大道235号恒运大厦3楼

Tel: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730



Test Report

Number: GZHT02340911

Tests Conducted (As Requested By The Applicant)

- 1 Water Resistance-Hydrostatic Pressure Test As Received (ANSI/AAMI PB 70:2012 /AATCC 127-2018, Option 2, Hydrostatic Head Tester, Pressure Gradient At 60 mbar/min Water Pressure):

Test On Fabric	Result			Requirement
	Area A (Front)	Area B (Sleeve)	Area C (Back)	
Specimen 1	164.0 cm	142.0 cm	163.0 cm	Level 3 >=50 cm
Specimen 2	162.0 cm	75.5 cm	170.5 cm	
Specimen 3	171.5 cm	144.0 cm	184.5 cm	

Test On Seam	Result				Requirement
	Seam At Shoulder	Seam At Armhole	Seam At Sleeve	Attachment Point	
Specimen 1	225.5 cm	180.5 cm	188.5 cm	187.5 cm	Level 3 >=50 cm
Specimen 2	236.5 cm	178.5 cm	182.5 cm	171.5 cm	
Specimen 3	173.5 cm	177.5 cm	171.5 cm	189.5 cm	

- 2 Water Resistance As Received -Impact Penetration Test Method (ANSI/AAMI PB 70:2012 /AATCC 42-2017):

Test On Fabric	Result			Requirement
	Area A (Front)	Area B (Sleeve)	Area C (Back)	
(Grams Of Water Absorbed)				Level 3 <=1.0g
Specimen 1	Negligible	Negligible	Negligible	
Specimen 2	Negligible	Negligible	Negligible	
Specimen 3	Negligible	Negligible	Negligible	

Test On Seam	Result				Requirement
	Seam At Shoulder	Seam At Armhole	Seam At Sleeve	Attachment Point	
(Grams Of Water Absorbed)					Level 3 <=1.0g
Specimen 1	Negligible	Negligible	Negligible	Negligible	
Specimen 2	Negligible	Negligible	Negligible	Negligible	
Specimen 3	Negligible	Negligible	Negligible	Negligible	

End of Report

This report is made solely on the basis of your instructions and/or information and materials supplied by you. It is not intended to be a recommendation for any particular course of action. Intertek does not accept a duty of care or any other responsibility to any person other than the Client in respect of this report and only accepts liability to the Client insofar as is expressly contained in the terms and conditions governing Intertek's provision of services to you. Intertek makes no warranties or representations either express or implied with respect to this report save as provided for in those terms and conditions. We have aimed to conduct the Review on a diligent and careful basis and we do not accept any liability to you for any loss arising out of or in connection with this report, in contract, tort, by statute or otherwise, except in the event of our gross negligence or wilful misconduct. No copy of the test report(except for full text copy) shall be made without the written approval by Intertek.

Ula / hilaryxu

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch

深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E801, No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China
 中国广州经济技术开发区科学城彩频路7号之二第1-8层02房、01房
 101、E201、E301、E401、E501、E601、E701、E801
 Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E801, No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China
 中国广州经济技术开发区科学城彩频路7号之二第1-8层02房、01房
 101、E201、E301、E401、E501、E601、E701、E801
 Tel: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730

